Mrs. 32 93 W 920 Dec 13

Not received yet.
29/10.
6/11
7/11
9 12/11
15/11
17/11

H 30 - H 31

Enclosure to P. W. Dept. Nos. 1048—62 M W, dated 3rd October 1879.

INVALID TABLE.

DR. HAMILTON'S PATTERN.

SPECIFICATION.

The drawing shows the full details of the table as suitable for the standard pattern, Broome's cot, which is 3′ 3″ in width from outer to outer of angle iron sides.

Enclosure 1 Tracing.

2. Figures 1 to 3 show the table with the revolving tray piece in a horizontal position, but figure 4 shows by dotted lines this tray in a sloping position and the mode of securing it at this slope with the iron forked pin. A small wooden cleat is fixed on the inside of each leg, so that the tray will rest on them when in an horizontal position.

3. Figures 5 and 6 show the details of the iron brackets, which, two on each leg, rest on the angle iron side pieces of the cot; the bottoms of these brackets are rounded to allow of the table riding more easily over the rivet heads in the angle iron side piece of the cot.

4. Figures 7 and 8 show the details of an iron pivot of the revolving tray and its bearings.

5. The wood work will be of well seasoned pine or other light kind of wood, and the weight of the complete article will not exceed 30 lbs.

6. The metal fittings will be of wrought iron, well finished and dammered.

J. W. WRIGHT, C. E.,
Officiating Superintending Engineer,
Sirhind Command.

No. 681, P. W. D. Press—16-10-79—50.

Nos. 1048—61 M W.

GOVERNMENT OF INDIA.

PUBLIC WORKS DEPARTMENT.

MILITARY WORKS.

To

The Governments of Madras, Bombay, Bengal, North Western Provinces and Oudh and Punjab, in the Public Works Department.

The Chief Commissioners of the Central Provinces, British Burmah, Mysore and Coorg, and Assam.

The Resident at Hyderabad.

The Agents of the Governor General for Central India and Rajpootana.

The Superintendent of Port Blair, and Chief Commissioner of the Andaman and Nicobar Islands.

The Inspector General of Military Works.

Simla, October 3, 1879.

Sir,

I am directed to forward herewith a plan, with specification, of Dr. Hamilton's pattern Invalid Table, and to request that it may be adopted in substitution for the present standard pattern bed table for future requirements.

I have the honor to be,
Sir,
Your most obedient Servant,

Captain, R. E.,

Asstt. Secy. to the Govt. of India.

No available spare copies of the letter & specification - 25 copies of plan enclosed.

P.W. 673/79

GOVERNMENT OF INDIA.

PUBLIC WORKS DEPARTMENT.

MILITARY WORKS.

To

Nos. 1048—62 M W.

Dated Simla, October 3, 1879.

Received

Forwards a plan and specification of Dr. Hamilton's pattern Invalid Table, and requests that it may be adopted in substitution for the present standard pattern bed table for future requirements.

No. 660, P. W. D. Press—8-10-79—50.

Enclosure to P. W. Dept. Nos. 1048—62 M W, dated 3rd October 1879.

INVALID TABLE.

DR. HAMILTON'S PATTERN.

SPECIFICATION.

Enclosure 1 Tracing.

The drawing shows the full details of the table as suitable for the standard pattern, Broome's cot, which is 3′ 3″ in width from outer to outer of angle iron sides.

2. Figures 1 to 3 show the table with the revolving tray piece in a horizontal position, but figure 4 shows by dotted lines this tray in a sloping position and the mode of securing it at this slope with the iron forked pin. A small wooden cleat is fixed on the inside of each leg, so that the tray will rest on them when in an horizontal position.

3. Figures 5 and 6 show the details of the iron brackets, which, two on each leg, rest on the angle iron side pieces of the cot; the bottoms of these brackets are rounded to allow of the table riding more easily over the rivet heads in the angle iron side piece of the cot.

4. Figures 7 and 8 show the details of an iron pivot of the revolving tray and its bearings.

5. The wood work will be of well seasoned pine or other light kind of wood, and the weight of the complete article will not exceed 30 lbs.

6. The metal fittings will be of wrought iron, well finished and dammered.

J. W. WRIGHT, C. E.,
Officiating Superintending Engineer,
Sirhind Command.


No. 681, P. W. D. Press—16-10-79—50.

INVALID TABLE

Dr. Hamilton's Pattern.

Fig. 7

DETAILS OF TABLE PIVOT AT G.G. *Fig 3*



B. Full sized Forked Pin to secure table in place

Fig. 8





Fig. 3


REAR ELEVATION ON A.B. *Fig 1*



Scale 1 foot to 6" for fig 5. 6. 7. 8.



Scale 1 to 1 for fig 1. 2. 3. 4.



Note The distance apart of side frames KK to be regulated according to the width of the cots for which the tables are required.

(Sd) James W. Wright C.E.
Offg Supdg Engr Sirhind Command
Military Works

Ordnance No 82935 of 20-12-79

Enclosure to P. W. Dept. Nos. 1048—62 M W, dated 3rd October 1879.

INVALID TABLE.

Dr. HAMILTON'S PATTERN.

Specification.

Enclosure 1 Tracing.

The drawing shows the full details of the table as suitable for the standard pattern, Broome's cot, which is 3′ 3″ in width from outer to outer of angle iron sides.

2. Figures 1 to 3 show the table with the revolving tray piece in a horizontal position, but figure 4 shows by dotted lines this tray in a sloping position and the mode of securing it at this slope with the iron forked pin. A small wooden cleat is fixed on the inside of each leg, so that the tray will rest on them when in an horizontal position.

3. Figures 5 and 6 show the details of the iron brackets, which, two on each leg, rest on the angle iron side pieces of the cot; the bottoms of these brackets are rounded to allow of the table riding more easily over the rivet heads in the angle iron side piece of the cot.

4. Figures 7 and 8 show the details of an iron pivot of the revolving tray and its bearings.

5. The wood work will be of well seasoned pine or other light kind of wood, and the weight of the complete article will not exceed 30 lbs.

6. The metal fittings will be of wrought iron, well finished and dammered.

J. W. WRIGHT, C. E.,
Officiating Superintending Engineer,
Sirhind Command.

No. 681, P. W. D. Press—16-10-79—50.

P.W.D.
C800178
No. 76 1/3

Mis No 3293 R of 20 December 79

Nos. 1048—61 M W.

GOVERNMENT OF INDIA.

PUBLIC WORKS DEPARTMENT.

MILITARY WORKS.

To

THE GOVERNMENTS OF MADRAS, BOMBAY, BENGAL, NORTH WESTERN PROVINCES AND OUDH AND PUNJAB, IN THE PUBLIC WORKS DEPARTMENT.

THE CHIEF COMMISSIONERS OF THE CENTRAL PROVINCES, BRITISH BURMAH, MYSORE AND COORG, AND ASSAM.

THE RESIDENT AT HYDERABAD.

THE AGENTS OF THE GOVERNOR GENERAL FOR CENTRAL INDIA AND RAJPOOTANA.

THE SUPERINTENDENT OF PORT BLAIR, AND CHIEF COMMISSIONER OF THE ANDAMAN AND NICOBAR ISLANDS.

THE INSPECTOR GENERAL OF MILITARY WORKS.

Simla, October 3, 1879.

SIR,

I am directed to forward herewith a plan, with specification, of Dr. Hamilton's pattern Invalid Table, and to request that it may be adopted in substitution for the present standard pattern bed table for future requirements.

I have the honor to be,
Sir,
Your most obedient Servant,

G. F. 2. Marshall Captain, R. E.,

Asstt. Secy. to the Govt. of India.

No. 1048 M W.
To The Secy to the Govt of Madras.
P.W. Department.

More Spare Copies will follow

I.W.D.
660/79
No. 7673

GOVERNMENT OF INDIA.

PUBLIC WORKS DEPARTMENT.

MILITARY WORKS.

To

Nos. 1048—62 M W.

Dated Simla, October 3, 1879.

Received

Mis No 32938 of 20-12-79

Forwards a plan and specification of Dr. Hamilton's pattern Invalid Table, and requests that it may be adopted in substitution for the present standard pattern bed table for future requirements.

No. 660, P. W. D. Press—8-10-79—50.

No 32938. P. W. Department.

PROCEEDINGS of the MADRAS GOVERNMENT.

Under date the 20th Dec 1879.

READ the following

From

The Govt of India — (Mily: 1048-62 M W dd Oct: 3rd 1879)

Communicated to Superintending Engineers.

29.12

To all Sirs

Mily. Dept.

Public Dept.

Dead 8-1-80

28.10